

99003026058000

Drinking water testing centres - Apply for approval

Heruntergeladen am 05.07.2025

<https://fimportal.de/xzufi-services/6006041-99003026058000/L100022>

Modul	Sachverhalt
Leistungsschlüssel	99003026058000
Leistungsbezeichnung I	Drinking water testing centres - Apply for approval
Leistungsbezeichnung II	Drinking water testing centres - Apply for approval
Typisierung	4 - Land: Regelung
Quellredaktion	Baden-Württemberg
Freigabestatus Katalog	unbestimmter Freigabestatus
Freigabestatus Bibliothek	unbestimmter Freigabestatus
Begriffe im Kontext	
Leistungstyp	
Leistungsgruppierung	
Verrichtungskennung	
SDG-Informationsbereich	
Lagen Portalverbund	
Einheitlicher Ansprechpartner	

Modul	Sachverhalt
Fachlich freigegeben am	
Fachlich freigegeben durch	
Handlungsgrundlage	Trinkwasserverordnung (TrinkwV) • §40 Zugelassene Untersuchungsstellen
Teaser	Drinking water tests required by the Drinking Water Ordinance (TrinkwV), including sampling, may only be carried out by authorised testing centres. With regard to the requirements for approval, the Drinking Water Ordinance of 23 June 2023 refers to the previously valid Drinking Water Ordinance (TrinkwV old version, Section 15 (4)).
Volltext	Drinking water tests required by the Drinking Water Ordinance (TrinkwV), including sampling, may only be carried out by authorised testing centres. With regard to the requirements for approval, the Drinking Water Ordinance of 23 June 2023 refers to the previously valid Drinking Water Ordinance (TrinkwV old version, Section 15 (4)).
Erforderliche Unterlagen	<ul style="list-style-type: none"> • Proof of valid accreditation as a testing laboratory for the analysis of drinking water, including sampling, • Confirmation of successful participation in round robin tests: for microbiological parameters, including legionella, at the Lower Saxony State Health Office, Aurich branch office, for chemical parameters at AQS Baden-Württemberg or Hamburg or a cooperation partner or at the North Rhine-Westphalia State Office for Nature, Environment and Consumer Protection The certificates must not be older than one year in each case. • when carrying out microbiological tests, confirmation of notification to the locally responsible regional council in accordance with Section 20 of the Federal Epidemics Act or Section 49 of the Protection against Infection Act and/or authorisation from the responsible regional council in accordance with Section 19 of the Federal Epidemics Act or Section 44 of the Protection against Infection Act.
Voraussetzungen	<ul style="list-style-type: none"> • a valid accreditation as a testing laboratory with the German Accreditation Body (DAkkS) or another

Modul	Sachverhalt
	<p>national accreditation body within the meaning of Regulation (EC) No. 765/2008 as amended for the performance of the required test procedures including sampling for the testing of drinking water in accordance with the TrinkwV,</p> <ul style="list-style-type: none"> • compliance with the requirements of Section 15 (1) to (2a) TrinkwV (old version) and • successful participation in external quality assurance programmes at least once a year.
Kosten	<p>Fees of EUR 350 to 500 are charged in accordance with the Ordinance of the Ministry of Rural Areas on the Determination of Fee Rates for Public Services of State Authorities in its Area of Responsibility (Fee Ordinance MLR - GebVO-MLR) of 11 December 2018.</p>
Verfahrensablauf	<p>The laboratory applies for authorisation informally, in writing or by e-mail, to the Baden-Württemberg Ministry of Food, Rural Areas and Consumer Protection.</p> <p>Enclose the required documents with the application.</p> <p>The Ministry checks the documents and decides on the authorisation. The laboratory will receive a notification. If approval is granted, the laboratory will be included in the list of approved testing centres published on the Ministry's website with its address, testing area and link to the accreditation body.</p>
Bearbeitungsdauer	2 to 4 weeks
Frist	none
weiterführende Informationen	
Hinweise	<p>Once accreditation has been granted, the drinking water testing centre must inform the Ministry of Food, Rural Areas and Consumer Protection of any changes to the accreditation immediately and without being asked in writing or by email. In addition, the drinking water testing laboratory must also submit confirmation to the Ministry by 31 December of each year that quality assurance has been successfully completed at least once a year by participating in interlaboratory</p>

Modul	Sachverhalt
	tests, also without being requested to do so. For microbiological parameters, including legionella, the Ministry expects participation in the round robin tests of the Lower Saxony State Health Office, Aurich branch, for chemical parameters at the AQS Baden-Württemberg or Hamburg or a cooperation partner or at the State Office for Nature, Environment and Consumer Protection North Rhine-Westphalia.
Rechtsbehelf	An appeal against the Ministry's decision may be lodged with the locally competent administrative court within the period specified in the decision.
Kurztext	
Ansprechpunkt	
Zuständige Stelle	
Formulare	
Ursprungsportal	